

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Addiese: COMMISSIONER FOR PATENTS P O Box 1450 Alexandra, Virginia 22313-1450 www.wepto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/634,368	08/06/2003	Kenneth Stewart	STEWART	6917
2644) 7590 977242908 BARNES & THORNBURG LLP 750-17TH STREET NW SUITE 900 WASHINGTON, DC 20006-4675			EXAMINER	
			SHAFFER, RICHARD R	
			ART UNIT	PAPER NUMBER
······································	1,502000 1075		3733	
			MAIL DATE	DELIVERY MODE
			07/24/2008	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

# Application No. Applicant(s) 10/634,368 STEWART, KENNETH Office Action Summary Examiner Art Unit Richard Shaffer 3733 -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS. WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). Status 1) Responsive to communication(s) filed on 22 May 2008. 2a) ☐ This action is FINAL. 2b) This action is non-final. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. Disposition of Claims 4) Claim(s) 23-39 is/are pending in the application. 4a) Of the above claim(s) 35 and 37 is/are withdrawn from consideration. 5) Claim(s) \_\_\_\_\_ is/are allowed. 6) Claim(s) 23-34,36,38 and 39 is/are rejected. 7) Claim(s) \_\_\_\_\_ is/are objected to. 8) Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement. Application Papers 9) The specification is objected to by the Examiner. 10) The drawing(s) filed on is/are; a) accepted or b) objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abevance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152. Priority under 35 U.S.C. § 119 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some \* c) None of: Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). \* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

Art Unit: 3733

### DETAILED ACTION

### Continued Examination Under 37 CFR 1.114

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on March 19<sup>th</sup>, 2008 has been entered.

#### Election/Restrictions

In the phone conversation with Evelyn Sommer on May 10<sup>th</sup>, 2006 and affirmed in the response filed on July 10<sup>th</sup>, 2006, the embodiment as shown in Figure 4 was elected without traverse. Newly added claims 35 and 37 are not directed towards the elected species of Figure 4. Claims 35 and 37 are withdrawn from further consideration pursuant to 37 CFR 1.142(b) as being drawn to a nonelected species, there being no allowable generic or linking claim.

### Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claim 29 is rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one

Art Unit: 3733

skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. Claim 29 recites the limitation "the hollow cap encapsulates substantially all of the bone instrumentation which projects from bone."

There is no disclosure or inherency as originally filed that the hollow cap encapsulates substantially all of the bone instrumentation which projects from bone.

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 26-28 and 34 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Claim 26 recites "a second suture string" with no appreciable claimed difference compared to a "plurality of first suture strings" as set forth in claim 25. Therefore, it is unclear where there are "second suture strings" with essentially the same properties as a "plurality of first suture strings." Claims 27 and 28 fail to correct for the deficiency.

Claim 34 recites the limitation "said hollow cap **or casing"** in line 2. There is insufficient antecedent basis for this limitation in the claim, specifically that of "or casing."

Art Unit: 3733

### Claim Rejections - 35 USC § 101

35 U.S.C. 101 reads as follows:

Whoever invents or discovers any new and useful process, machine, manufacture, or composition of matter, or any new and useful improvement thereof, may obtain a patent therefor, subject to the conditions and requirements of this title.

Claims 23-34, 38 and 39 are rejected under 35 U.S.C. 101 because the claimed invention is directed to non-statutory subject matter. As previously conveyed to applicant in the interview on May 21<sup>st</sup>, 2008, applicant cannot positively recite a portion of the human body. In the claims rejected under this statute, applicant has positively recited the cover being installed over bone instrumentation which has been placed within bone. For examination purposes, the claims will be interpreted as being a bone instrumentation cover capable of being placed over bone instrumentation during bone repair consistent with claim 36 which is not rejected under this statute.

### Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 23-34, 36, 38 and 39 are rejected under 35 U.S.C. 102(b) as being anticipated by Ellman (US Patent 4,428,375).

Ellman discloses a device (**Figures 1-4**) comprising: a hollow cap (**15**) made of absorbable Dacron<sup>™</sup> (**Column 2, Lines 30-40**) and capable of removal in a subsequent surgical procedure; a separate tubular cinch ring (**20 and 22 together**) having openings

Art Unit: 3733

(one opening defined by the enclosed circles of either 20 or 22; a second opening defined by the space between 20 and 22 – same opening direction as those of 19); adjustable sutures (16) of the cap (15) are "threaded" through the openings of the cinch ring; [First Interpretation] a non-adjustable (as in same direction as adjustable) suture (21) tied to the lower edge of the cap and attached to the cinch ring; [Second Interpretation] a drawstring (21) within the cinch ring to circumferentially tighten the hollow cap.

## Response to Arguments

Applicant's arguments with respect to the claims have been considered but are moot in view of the new ground(s) of rejection.

#### Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Richard Shaffer whose telephone number is (571)272-8683. The examiner can normally be reached on Monday-Friday (7am-5pm).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/634,368 Page 6

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Richard Shaffer/ Examiner, Art Unit 3733 /Eduardo C. Robert/ Supervisory Patent Examiner, Art Unit 3733